

# sifin

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 927 030-0  
F: +49 30 927 030-30  
info@sifin.de  
www.sifin.de

<b>Product</b>	<b>Salmonella Typhi-OH-Test Antigen</b> (9,12:d:-) for the Widal Reaction
Catalogue no.	TS 1616
Package size	10 ml
<b>Lot no.</b>	<b>170 08 15</b>
<b>Expiry date</b>	<b>2016-08</b>
Intended use	The Test antigen reacts in the Widal test with the homologous <i>Salmonella</i> antibodies (Anti O:9,12; Anti-H:d) in the serum to be tested. The specific bonding is indicated by agglutination.
Specificity	There must be no reaction with physiological saline solution.
Preservatives	Formalin 0.5 %
Storage	2 °C...8 °C

Berlin, 01.09.15  
\_\_\_\_\_  
place, date

  
\_\_\_\_\_  
signature Head of Quality Control