

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Control Serum Anti-Salmonella H:b for the Widal Reaction
Catalogue no.	TS 1642
Package size	1 ml
Lot no.	3570417
Expiry date	2020-04
Intended use	The control serum agglutinates in the Widal test with the homologous Salmonella H:bTest Antigen (sifin) with a titre of 1:1600 (tube). Deviations of \pm one titre step are permitted.
Specificity	The serum control (antigen addition is substituted by phosphate buffered saline solution) must not show any agglutination.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin,
place, date

9.05.2017

signature Head of Quality Control

