

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Shigella dysenteriae type 2
Catalogue no.	TS 2002
Package size	1 ml
Lot no.	0160119
Expiry date	2024-01
Intended use	Anti-Shigella dysenteriae type 2 records all <i>Shigella dysenteriae</i> strains, as long as they contain the serotype 2 in slide agglutination.
Specificity	No agglutination with serotypes of the serotypes of <i>Shigella dysenteriae</i> types 1, 3 to 10, <i>Shigella flexneri</i> , <i>Shigella boydii</i> and <i>Shigella sonnei</i> (phase I and phase II).
Preservatives	Sodium azide 0.9 mg / ml
Storage	2 °C...8 °C

Berlin,
place, date

25.07.2019

signature Head of Quality Control

