



BIOMEDICAL SYSTEMS

370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.elitechgroup.com

2021, June 22

Certificate of Analysis:

Standard Solution Ampule, ChloroChek®

Catalog Number: SS-251
Description: Standard Solution Ampule (100 mmol/L), ChloroChek®
Lot: 20384051
Expiration Date: 2025, September 16

This is to certify that ELITechGroup Biomedical Systems ChloroChek Standard Solution noted above was tested by Gonotec for ELITechGroup Biomedical Systems. The standard solution is traceable to standards prepared from *NIST standard reference material and was determined to meet required specifications. Results are in the table below.

Test Results		
Characteristic	Requirement	Result
Ampule color	Clear glass	Colorless glass
Appearance of solution	Clear and colorless solution, practically free of visible particles	Passes test
pH	5.0 – 7.0	7.0
Extractable volume	≥ 1.0 ml	1.1 ml
Chloride	99.0 – 101.0 mmol/l	99.7 mmol/l

Authorized by:

Whitney Lewis, Director of Quality

*NIST refers to United States Department of Commerce National Institute of Standards and Technology, located in Gaithersburg, Maryland, USA.