

Quality control certificate

Mycobiotic Agar (Fungal Selective Agar)

Nº Cat: 1072

Nº Batch: 301042

Re-test Date: 1/2026



Formula in g/l

Bacteriological agar	15,5	Chloramphenicol	0,05
Cycloheximide	0,4	Dextrose	10
Soy peptone	10		

Typical formula g/L * Adjusted and/or supplemented as required to meet performance criteria.

Quality control

Solubility	Appearance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Batch pH
w/o rests	Fine powder	Beige	Amber, slightly opalescent	6,9±0,2	6,9

Microbiological test

Incubation conditions: (22-25 °C, 35 °C / 3-7 days).

Microorganisms	Specification
Aspergillus brasiliensis ATCC 16404	Total-partial inhibition
Candida albicans ATCC 2091	Good growth
Trichophyton mentagrophytes ATCC 24957	Buen crecimiento
Escherichia coli ATCC 25922	Total inhibition
Staphylococcus aureus ATCC 25923	Total inhibition
Trichophyton rubrum CECT 2794	Good growth

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.