



BIOMEDICAL SYSTEMS  
 370 West 1700 South  
 Logan, Utah 84321 USA  
 435 752 6011  
 800 453 2725  
 fax 435 752 4127  
 web www.elitechgroup.com

2021, August 31

**Certificate of Analysis:**

Macroduct® Advanced Supply Kit

**Catalog Number:** SS-268  
**Lot for SS-268** 226902  
**Expiration Date:** 2023, April 30  
**Description:** Pilogel Discs for Macroduct® Advanced Supply Kit  
**Lot for 32-0189-01:** 226956  
**Expiration Date:** 2023, April 30

This is to certify that the product noted above is manufactured under ISO 13485:2016, and FDA 21 CFR 820 Standards. Macroduct® Advanced Supply Kit and its packaging are inspected for damage or defects in material and workmanship. All tests were performed using samples of the batch product. All test instruments are calibrated annually and verified before use with NIST\* traceable standards.

Test Results		
Test Performed	Specification/Reference	Result
Weight to Disc	4.292 to 5.140 (g)	4.800
Corrected Absorbance A <sub>206</sub> for Weight	1.469 to 1.691	1.556
Color Verified	Semitransparent in appearance with slight yellow to brown	Pass
Texture Verified	Cracks in or on the surface of pilogel disc unacceptable	Pass
	Perimeter of the disc, smooth	Pass
	Ends of the disc smooth to slightly wavy	Pass
	Debris of any kind is not permitted	Pass

Authorized by:

 2021/08/31  
 Whitney Lewis, Director of Quality

\*NIST refers to United States Department of Commerce National Institute of Standards and Technology, located in Gaithersburg, Maryland, USA.