



BIOMEDICAL SYSTEMS
 370 West 1700 South
 Logan, Utah 84321 USA
 435 752 6011
 800 453 2725
 fax 435 752 4127
 web www.elitechgroup.com

2021, November 15

Certificate of Analysis:

Macroduct® Advanced Supply Kit

Catalog Number: SS-268
Lot for SS-268 227094
Expiration Date: 2023, April 30
Description: Pilogel Discs for Macroduct® Advanced Supply Kit
Lot for 32-0189-01: 226958
Expiration Date: 2023, April 30

This is to certify that the product noted above is manufactured under ISO 13485:2016, and FDA 21 CFR 820 Standards. Macroduct® Advanced Supply Kit and its packaging are inspected for damage or defects in material and workmanship. All tests were performed using samples of the batch product. All test instruments are calibrated annually and verified before use with NIST* traceable standards.

Test Results		
Test Performed	Specification/Reference	Result
Weight to Disc	4.292 to 5.140 (g)	4.852
Corrected Absorbance A ₂₀₆ for Weight	1.469 to 1.691	1.5811
Color Verified	Semitransparent in appearance with slight yellow to brown	Pass
Texture Verified	Cracks in or on the surface of pilogel disc unacceptable	Pass
	Perimeter of the disc, smooth	Pass
	Ends of the disc smooth to slightly wavy	Pass
	Debris of any kind is not permitted	Pass

Authorized by:  01/12/2022

Lyndon Jay Smith, Quality Assurance Associate

*NIST refers to United States Department of Commerce National Institute of Standards and Technology, located in Gaithersburg, Maryland, USA.