



CERTIFICATE OF ANALYSIS

ESR_COA_SI305#2432A.pdf

Product Code SI 305.300-A	Name LATEX CONTROLS	Quantity manufactured 178	Type of Tubes Greiner
-------------------------------------	-------------------------------	-------------------------------------	---------------------------------

Lot Number 2432A	Manufacturing Date 14/09/2023	Expiry Date 30/12/2023		
----------------------------	---	----------------------------------	--	--

Standard Values			Measured Values				Internal	Volume
Transmittance	ESR (mm/h) (min-Val-max)		ESR (mm/h) (min-Val-max)		Instrument	Serial Number	Aggregation	checked
Level 2	1534	6 - 11	9 9 9	TEST1 THL	6608	<input checked="" type="checkbox"/>	NO	YES
Level 3	980	15 - 22	19 20 20	TEST1 THL	T13290	<input checked="" type="checkbox"/>	NO	YES
Level 4	637	56 - 74	65 66 68	ROLLER 20 PN	133	<input checked="" type="checkbox"/>	NO	YES

Controlled By **Cirio Alberto**
Approved By Manufacturing Manager **Pividori Denis**

Date of Control **14/09/2023**
Date of Approval **14/09/2023**

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance **Trevisan Antonia**

Date of Release **14/09/2023**

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
Tel (+39) 049 0992000 - Fax (+39) 049 5855434
Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
Tel (+39) 0432 547454 - Fax (+39) 0432 547378
Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com