

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

Product	Anti-Shigella III Group C S. boydii type 1-15
Catalogue no.	TS1831
Package size	1 ml
Lot no.	9541222
Expiry date	2027-11
Intended use	Anti-Shigella III records the serotypes of <i>Shigella boydii</i> (type 1 to 15) in slide agglutination.
Specificity	No agglutination with the serotypes of <i>Shigella flexneri</i> , <i>Shigella dysenteriae</i> (type 1 to 15) and <i>Shigella sonnei</i> (Phase I and Phase II).
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

**This document has been established electronically and is valid without signature.
The batch release was done in accordance with SOP QMV 7.**