# Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

**sifin diagnostics gmbh**

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 700 144-0
F: +49 30 700 144-30
info@sifin.de
www.sifin.de

| | |
|---|---|
| **Product** | **Anti-Salmonella H:f** |
| Catalogue no. | TR1407 |
| Package size | 1 ml |
| **Lot no.** | **2730423** |
| **Expiry date** | **2028-03** |
| Intended use | Anti-Salmonella H:f records all *Salmonella* strains, as long as they contain in an antigen combination the H f antigen (e. g. H:f,g; H:f,g,s; H:f,g,t; H:f,g,m,t). In slide agglutination. |
| Specificity | No agglutination with heterologous *Salmonella* H antigens and *Salmonella* O antigens. |
| Preservatives | Sodium azide 0.9 mg/ml |
| Storage | 2 °C…8 °C |

**This document has been established electronically and is valid without signature.**
**The batch release was done in accordance with SOP QMV 7.**