



## EC DECLARATION OF CONFORMITY

MANUFACTURER:	DIESSSE DIAGNOSTICA SENESE SPA STRADA DEI LAGHI 39 53035 MONTERIGGIONI (SI), ITALY
EUROPEAN REPRESENTATIVE:	//
GENERIC NAME	AUTOMATIC INSTRUMENT FOR ESR ANALYSIS
PRODUCT: CODE:	<b>MINI-CUBE</b> <b>10392</b>
TECHNICAL DATA:	9V DC; 2 A
CLASSIFICATION:	IVD NOT IN ANNEX II OR SELF-TESTING IVD
CONFORMITY ASSESSMENT ROUTE:	ANNEX APPLIED N° III EXCLUDING (6) ESSENTIAL REQUIREMENTS OF ANNEX I

WE HEREWITH DECLARE THAT THE DESIGN, TYPE OF MANUFACTURE OF THE IN VITRO MEDICAL DIAGNOSTIC DEVICE DESCRIBED ABOVE AND THE VERSION DISTRIBUTED ON THE MARKET, CONFORMS TO THE 98/79/EEC DIRECTIVE RELEVANT TO THE IN VITRO MEDICAL-DIAGNOSTICS DEVICES (IVD)

### THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

### THE PRODUCT CONFORMS, AS A WHOLE AND IN ITS PARTS, WITH THE FOLLOWING STANDARDS AND THEIR AMENDMENTS:

EN 61010-1:2010	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 1: GENERAL REQUIREMENTS
EN 61010-2-081:2015	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-081: PARTICULAR REQUIREMENTS FOR AUTOMATIC AND SEMI-AUTOMATIC LABORATORY EQUIPMENT FOR ANALYSIS AND OTHER PURPOSES
EN 61010-2-101:2002	SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-101: SAFETY REQUIREMENTS FOR IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

EN 61326-1:2013 ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 1: GENERAL REQUIREMENTS

EN 61326-2-6:2013 ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 2-6: PARTICULAR REQUIREMENTS - IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

**AND THEREFORE MEETS THE ESSENTIAL REQUIREMENTS OF THE FOLLOWING COMMUNITY DIRECTIVES AND THEIR AMENDMENTS:**

LOW VOLTAGE DIRECTIVE (2014/35/EU)

ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)

RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT DIRECTIVE – ROHS2 (2011/65/EU)

NOTIFIED BODY: NOT NECESSARY

(EC) CERTIFICATE: N.A.

START OF CE-MARKING: JUNE 2016

REVISION: 8

PLACE, DATE OF ISSUE: MONTERIGGIONI, 25 MAY 2022

EXPIRY DATE: 25 MAY 2027

**THE PRESENT EC DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.**

SIGNATURE:



\_\_\_\_\_  
CHIARA MUZZI  
REGULATORY AFFAIRS MANAGER

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI, 25/05/2022



\_\_\_\_\_  
MAGDALENA STOCZKO  
REGULATORY SUPERVISOR