

• Certificate of Analysis

Product Name : **MBT STAR-Carba IVD Kit**
 Part No.: **1848467**
 LOT No.: **0000484352**
 Use by¹⁾: **2025-02-28**
 Date of Production: **2024-03-20**
 Storage Conditions: **The MBT STAR-Carba IVD Kit is shipped at ambient temperature but must be stored at -18 to -25°C immediately on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
1. Instrument calibration for MBT STAR-BL assay mass range using MBT STAR-BL IVD Module.	Successfull automated MBT STAR Calibrator based instrument calibration.	Passed
2. MBT STAR-BL assay functional performance test using MBT STAR-Carba IVD Kit in accordance with the IFU.	MBT STAR-BL IVD Module results: <ol style="list-style-type: none"> negative control (- ctrl) as non-hydrolyzed positive control (+ ctrl; KPC) as hydrolyzed negative QC Sample (- QC) as non-hydrolyzed positive QC Sample (+ QC; NDM-1) as hydrolyzed 	Passed Passed Passed Passed


Remarks:

Kit components have to be handled carefully to avoid contaminations.

¹⁾ If stored unopened at -18 to -25°C the product is stable until expiration date.

Bruker Daltonics GmbH & Co. KG, Germany
 Fahrenheitstr. 4 – 28359 Bremen
 phone +49-421-2205-0
 QC release date: 2024-03-20




 Bruker Daltonics GmbH & Co. KG
 Fahrenheitstr. 4
 D-28359 Bremen
 Phone +49 (0)421-2205-0
 Fax +49 (0)421-2205-370
Dr. Thomas Maier, SAFETY OFFICER