


### CERTIFICATE OF ANALYSIS

<b>Product Name: Prolex™ Staph Xtra Latex Kit</b>		<b>Lot Number: C22249</b>
<b>Product Code: PL.1080</b>		<b>Expiration Date: 2025-07-31</b>
<b>Product Code</b>	<b>Quantity</b>	<b>Components</b>
PL.1083	2 x 2.5 ml	Staph Xtra Latex Reagent
PL.1085	1 x 5.0 ml	Staph Xtra Negative Control Latex Reagent
	200	Mixing Sticks
	20 cards	Test Cards
<b>Additional Information</b>		
Storage Conditions:		2-8°C
Instructions for Use:		Refer to pro-lab.com/resources for current version
<b>Quality Control Testing</b>		
<b>Criteria</b>		<b>Result</b>
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.		PASS
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.		PASS
* MRSA - (Methicillin-resistant <i>Staphylococcus aureus</i> )		
<b>Release</b>		
2023-12-19		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424  
 21 Cypress Blvd., Suite 1155, Round Rock,  
 Texas 78665-1046  
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.co.uk