



# CERTIFICATE OF ANALYSIS

MIC\_COA\_SI390900#SCR24111.pdf

Product Code <b>SI 390.900</b>	Name <b>URO-QUICK SCREENING KIT</b>	Quantity manufactured <b>27</b>	
-----------------------------------	--	------------------------------------	--

Lot Number <b>SCR24111</b>	Manufacturing Date <b>30/04/2024</b>	Expiry Date <b>29/06/2025</b>	CAP color <b>Green</b>	Printed Band <b>White</b>
-------------------------------	---	----------------------------------	---------------------------	------------------------------

## Eugonic Broth

Lot Number **SCR24111**  
Expiry Date **29/06/2025**  
CAP color **Green**  
Printed Band **White**

	Physical parameters (Volume, transparency, colour)	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% ≤ 35%	CV% ≤ 35%
Found	OK	OK	14.0%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Controlled By ..... **Cirio Alberto**

Date of Control **06/05/2024**

Approved By Manufacturing Manager **Pividori Denis**

Date of Approval **06/05/2024**

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance ..... **Gaddi Giulia**

Date of Release **06/05/2024**

This is a computer-generated document. No signature is required.

### Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy  
Tel (+39) 049 0992000 - Fax (+39) 049 5855434  
Mail PEC: info.alifax@pec.alifax.com  
[info@alifax.com](mailto:info@alifax.com) - [www.alifax.com](http://www.alifax.com)



### Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy  
Tel (+39) 0432 547454 - Fax (+39) 0432 547378  
Mail PEC: info.udine@pec.alifax.com  
[info.udine@alifax.com](mailto:info.udine@alifax.com)