

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

|                    |   |
|--------------------|---|
| <b>Product</b>     | <b>Anti-Shigella III</b><br>S. boydii type 1-15   |
| Catalogue no.      | TS1831  |
| Package size       | 1 ml  |
| <b>Lot no.</b>     | <b>5700723</b>  |
| <b>Expiry date</b> | <b>2028-06</b>  |
| Intended use       | Anti-Shigella III records the serotypes of <i>Shigella boydii</i> (type 1 to 15) in slide agglutination.  |
| Specificity        | No agglutination with the serotypes of <i>Shigella flexneri</i> , <i>Shigella dysenteriae</i> (type 1 to 15) and <i>Shigella sonnei</i> (Phase I and Phase II). |
| Preservatives      | Sodium azide 0.9 mg/ml  |
| Storage            | 2 °C...8 °C   |

**This document has been established electronically and is valid without signature.  
The batch release was done in accordance with SOP QMV 7.**