

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

Product	Anti-Coli O 145:K-
Catalogue no.	TS2221
Package size	1 ml
Lot no.	8981123
Expiry date	2028-10
Intended use	Anti-Coli O 145:K- records all <i>E. coli</i> strains that contain the antigen O 145:K- in slide agglutination. In the Widal reaction the serum reaches a titre of 1:1600 (tube and MTP) with the homologous <i>E. coli</i> strains; deviations of ± 1 titre step are acceptable.
Specificity	Since unspecific co-agglutinations with other <i>E. coli</i> strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled <i>E. coli</i> isolate in the Widal test.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2 °C...8 °C

**This document has been established electronically and is valid without signature.
The batch release was done in accordance with SOP QMV 7.**