

CERTIFICATE OF ANALYSIS

Product Name:	lame: Prolex™ Staph Xtra Latex Kit			C22638	
Product Code:	PL.1080		Expiration Date:	2025-10-31	
Product Code	Quantity	Compone	Components		
PL.1083	2 x 2.5 ml	Staph Xtra	aph Xtra Latex Reagent		
PL.1085	1 x 5.0 ml	Staph Xtra			
	200	Mixing Stic	ks		
	20 cards	Test Cards			
Additional Inform					
Storage Conditions: 2-8°C					
Instructions for Use: Refer to pro-lab.com/resources for current version					
Quality Control	Testina				
Criteria				Result	
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.				PASS	
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.				PASS	
* MRSA - (Methic	illin-resistant <i>Stap</i>	hylococcus aureu	s)		
Release					
2024-04-16			2		
Date			Quality Control		