

QUALITY CONTROL CERTIFICATE

Posaconazole 5 µg Neo-Sensitabs

REF: 82612

LOT: 24-1030

Date of analysis: 2024-10

Date of expiry: 2029-10

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:
C. albicans ATCC 90028
C. krusei ATCC 6258
C. parapsilosis ATCC 22019

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mea
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality