



www.pro-lab.co.uk

Certificate of analysis

Product Name: GRAMS DIFFERENTIATOR	Lot number: K38257
Product Code: PL.7006	Expiry date: 2026 06 30

Additional Information

Description: Clear colourless liquid.	Concentrate Lot number: n/a Production date: 2024/06/03 Instructions for Use:	Storage conditions: 15-25°C Precautions: refer to SDS Go to pro-lab.co.uk for current version.
---------------------------------------	---	---

Quality Control Testing

Challenged with test organisms:
Escherichia coli NCTC 12241
Oxford Staphylococcus aureus NCTC 12981

Test results:
Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:
We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2024-06-05	
Date	Signed

T: 0151 353 1613
3 Bassendale Road, Bromborough,
Wirral, CH62 3QL, UK.
uksupport@pro-lab.com