

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert® X	Apress Flu/RSV		
Cepheid Catalogue Part No.:	XPRSFLU/RSV-CE-10		
Kit Lot No.: 1001440689			
Cartridge Lot No.: 46410			
Kit Expiration Date: 2026-07-	19		
Legal Manufacturer	Manufacturing Facility	Solna	Sunnyvale
Cepheid		Newark	Lodi
904 Caribbean Drive			
Sunnyvale, CA 94089			
USA			

Funct	ional	Testing
		2 00

Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

	MAN	08/02/2021
Signati	re of Quality Assurance,	Date
Name:	Jennifer Nguyen	He m
Title:	Quality Assurance Supervis	or