

CERTIFICATE OF ANALYSIS

| | |
|--------------------------------------------------------------------|------------------------------------|
| Product Name: Streptococcal Grouping - Extraction Reagent 3 | Lot Number: C22758B |
| Product Code: PL.039 | Expiration Date: 2026-05-31 |

Additional Information


| | |
|-----------------------|----------------------------------------------------|
| Storage Conditions: | 2-8°C |
| Instructions for Use: | Refer to pro-lab.com/resources for current version |

Quality Control Testing

| Organism | Reaction* | Result |
|-------------------------------------------------------|-----------|--------|
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A) | Positive | PASS |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B) | Positive | PASS |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Positive | PASS |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D) | Positive | PASS |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F) | Positive | PASS |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G) | Positive | PASS |

*When used in conjunction with Extraction Reagents 1 and 2 to extract group specific carbohydrate antigens and tested against homologous latex reagents.

Release

| | |
|------------|--------------------------------------------------------------------------------------|
| 2024-07-05 |  |
| Date | Quality Control |