



www.pro-lab.co.uk

Certificate of analysis

Product Name: CRYSTAL VIOLET	Lot number: K38492
Product Code: PL.7000	Expiry date: 2025 07 31

Additional Information

Description: Deep violet solution with slight odour.	Concentrate Lot number: K38450 Production date: 2024/07/05 Instructions for Use:	Storage conditions: 15-25°C Precautions: refer to SDS Go to pro-lab.co.uk for current version.
--	--	---

Quality Control Testing

Challenged with test organisms:
Escherichia coli NCTC 12241
Oxford Staphylococcus aureus NCTC 12981

Test results:
Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:
We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2024-07-09	
Date	Signed

T: 0151 353 1613
3 Bassendale Road, Bromborough,
Wirral, CH62 3QL, UK.
uksupport@pro-lab.com