

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR) and the China National Medical Products Administration's Quality System Requirements.

**Product Name:** 

Xpert® HBV Viral Load

Cepheid Catalogue Part No.: GXHBV-VL-CE-10

Kit Lot No.:

1001444804

Cartridge Lot No.:

30202

**Kit Expiration Date:** 

2026-02-01

Legal Manufacturer

**Manufacturing Facility** 

Solna Sunnyvale

Cepheid AB

Röntgenvägen 5

SE-17154 Solna

Sweden

Cepheid AB

Röntgenvägen 5 SE-171 54 Solna

Sweden

## Functional Testing according to D51114 Rev: D

Test Description	Acceptance Criteria	Test Result
LOW Positive	HBV DETECTED: 1.26-1.70 log10 IU/mL	Passed
HIGH Positive	HBV DETECTED: 5.54-5.86 log10 IU/mL	Passed
Negative	HBV NOT DETECTED	Passed

If checked this document is produced electronically and valid without a wet signature.

Signature of Quality Assurance

Name: Robert Fiedler

Title: QA Analyst