

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert	® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part No	o.: XP3COV2/FLU/RSV-	10	
4004404047			
Kit Lot No.: 1001104617			
Cartridge Lot No.: 16402			
Kit Expiration Date: 2024-0	07-28		
Legal Manufacturer	Manufacturing Facility	Solna	Sunnyvale
Cepheid	Cepheid AB	Newark	Lodi IVD (B2)
904 Caribbean Drive	Röntgenvägen 5		
Sunnyvale, CA 94089	SE-171 54 Solna		
LISA	Sweden		

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

☐ If checked this document is produced electronically and valid without a wet signature.

1	L 2023-0	08-06
Signat	ure of Quality Assurance,	Date
Name:	Alexander Avramidis	
Title:	QA Analyst	