

Certificate of Analysis

Page: 1 of

2

Becton Dickinson Caribe LTD. BD Diagnostics Systems 6 Vicks Drive, Lot # 6 Cayey PR 00737-2860 PR

Product Name : Mycosis I/Cbactec F 50/Pk F/G

Catalog Number : 442206 Manufacture Date: 2021/04/13

Batch Number : 1097816 Expiration Date : 2022/07/31

This is to certify that representative samples of

BACTEC MYCOSIS-IC/F MEDIUM

were tested in the Quality Control Laboratory by procedures conventionally utilized for this type of product and met the following test parameters:

pH: 6.8 ± 0.2

Autoclaving:

The product was exposed to a moist heat sterilization process (previously validated following an ISO Standard*).

Vacuum draw:

greater than or equal to 10 ml at time of manufacture.

Biological Performance: Satisfactory growth:

| CULTURE | ATCC No |
|-------------------------|---------|
| Candida albicans | 10231 |
| C. albicans | 14053 |
| C. tropicalis | 750 |
| C. parapsilosis | 10232 |
| C. krusei | 34135 |
| Torulopsis glabrata | 15545 |
| Cryptococcus neoformans | 13690 |

Bacterial Inhibition:

No growth:

Escherichia coli 25922 Staphylococcus aureus 25923

ATCC is a trademark of the American Type Culture Collection.

* ISO 11134, Sterilization of health care products - Requirements for validation and routine control - Industrial moist heat sterilization, 1994.

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems is an ISO 13485:2016 Registered facility. BD Diagnostics - Diagnostic Systems products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BD Diagnostics - Diagnostic

Creation Date: 2021/04/30 19:33:24



Certificate of Analysis

Page: 2 of

Becton Dickinson Caribe LTD.
BD Diagnostics Systems
6 Vicks Drive, Lot # 6
Cayey PR 00737-2860 PR

Product Name : Mycosis I/Cbactec F 50/Pk F/G

Catalog Number : 442206 Manufacture Date: 2021/04/13

Batch Number : 1097816 Expiration Date : 2022/07/31

Systems stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see

www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information.

2)gu-

Zuleika Vargas Senior Quality

Senior Quality Manager Signature Date: 2021/04/30

Creation Date: 2021/04/30 19:33:24