

Certificate of Analysis

Becton Dickinson Caribe LTD. BD Diagnostics Systems 6 Vicks Drive, Lot # 6 Cayey PR 00737-2860 PR Page: 1 of 2
Product Name: Bactec Plastic Mycosis Ic/F MediumCatalog Number: 442017Manufacture Date: 2022/05/07Batch Number: 2116365Expiration Date: 2023/08/31
This is to certify that representative samples of BACTEC MYCOSIS-IC/F MEDIUM were tested in the Quality Control Laboratory by procedures conventionally utilized for this type of product and met the following test parameters:
pH: 6.8 ± 0.2
Autoclaving: The product was exposed to a moist heat sterilization process (previously validated following an ISO Standard*).
Vacuum draw: greater than or equal to 14 ml at time of manufacture.
Biological Performance: Satisfactory growth:
CULTUREATCC No.C. aurisCDC-AR-0387C. albicans14053C. tropicalis750C. parapsilosis10232C. krusei34135C. glabrata15545Cryptococcus neoformans13690
Bacterial Inhibition: No growth: Escherichia coli 25922 Staphylococcus aureus 25923
ATCC is a trademark of the American Type Culture Collection. * ISO 17665, Sterilization of health care products - Moist heat - Requirements for the development, validation and routine control of a sterilization process for Medical Devices.
The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.
BD Diagnostic Systems is an ISO 13485:2016 Registered facility. BD Diagnostics - Diagnostic Systems products are manufactured in facilities registered with the United States Food and Drug



Becton Dickinson Caribe LTD. BD Diagnostics Systems 6 Vicks Drive, Lot # 6 Cayey PR 00737-2860 PR

Page: 2 of 2

Product Name	: Bactec Plastic Mycosis Ic/F Medium	
Catalog Number	: 442017 Manufacture Date: 2022/05/07	
Batch Number	: 2116365	
Expiration Date	: 2023/08/31	

Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BD Diagnostics - Diagnostic Systems stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see

www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customerregulatory-support-information.

Zuleika Vargas Senior Quality Manager Signature Date:2022/05/23