

## Certificate of Analysis

2

Page: 1 of

Becton Dickinson Caribe LTD. BD Diagnostics Systems 6 Vicks Drive, Lot # 6 Cayey PR 00737-2860 PR

Product Name : Bactec Plastic Mycosis Ic/F Medium

Catalog Number : 442017 Manufacture Date: 2023/09/19

Batch Number : 3256411 Expiration Date : 2024/12/17

This is to certify that representative samples of

BACTEC MYCOSIS-IC/F MEDIUM

were tested in the Quality Control Laboratory by procedures

conventionally utilized for this type of product and met the following test parameters:

pH: 6.8 ± 0.2

Autoclaving:

The product was exposed to a moist heat sterilization process (previously validated following an ISO Standard\*).

Vacuum draw:

greater than or equal to 14 ml at time of manufacture.

Biological Performance: Satisfactory growth:

CULTURE ATCC No. C. auris CDC-AR-0387 C. albicans 14053 C. tropicalis 750 C. parapsilosis 10232 C. krusei 34135 C. glabrata 15545 Cryptococcus neoformans 13690

Bacterial Inhibition:

No growth:

Escherichia coli 25922 Staphylococcus aureus 25923

ATCC is a trademark of the American Type Culture Collection.

\* ISO 17665, Sterilization of health care products - Moist heat - Requirements for the development, validation and routine control of a sterilization process for Medical Devices.

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems is an ISO 13485:2016 Registered facility. BD Diagnostics - Diagnostic Systems products are manufactured in facilities registered with the United States Food and Drug

Creation Date: 2023/10/06 19:42:38



## **Certificate of Analysis**

Page: 2 of

Becton Dickinson Caribe LTD.
BD Diagnostics Systems
6 Vicks Drive, Lot # 6
Cayey PR 00737-2860 PR

Product Name : Bactec Plastic Mycosis Ic/F Medium

Catalog Number : 442017 Manufacture Date: 2023/09/19

Batch Number : 3256411
Expiration Date : 2024/12/17

Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BD Diagnostics - Diagnostic Systems stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see

www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information.

2) gre-

Zuleika Vargas Director, Quality Management Signature Date: 2023/10/06

Creation Date: 2023/10/06 19:42:38