

# • Certificate of Analysis

**Product Name :** MBT Sepsityper IVD Kit  
**Part No.:** 1834338  
**LOT No.:** 6010524011  
**Use by<sup>1)</sup>:** 2026-08-31  
**Date of Production:** 2024-10-25  
**Storage Conditions :** All kit components must be stored at room temperature (15–25 °C).

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Identification of <i>Candida glabrata</i> from a blood culture using the MALDI Biotyper and the MBT Sepsityper IVD Kit following the kit sample preparation procedure.	Classification Result Report lists „ <i>Candida glabrata</i> ” as the best match.	Passed
	log(score) Value $\geq 2.00$ of the best match.	Passed
Identification of blank test samples using the MALDI Biotyper and the MBT Sepsityper IVD Kit following the kit sample preparation procedure.	Classification Result Report lists „no peaks found” or “No Organism Identification Possible” as the best match.	Passed
	log(score) Value $< 1.60$ of the best match.	Passed

**Remarks:**

<sup>1)</sup> Kit components should be handled carefully to avoid contaminations.

Bruker Daltonics GmbH & Co. KG  
 Fahrenheitstr. 4  
 28359 Bremen  
 Germany  
 Phone +49-421-2205-0

  
 Quality Management

QC release date: 2024-10-28