

CERTIFICATE OF ANALYSIS

Product Name: Streptococcal Grouping - Prolex™ Extraction Reagent Set	Lot Number: C20094
Product Code: PL.046	Expiration Date: 2023-10-31

Product Code	Quantity	Components
PL.047	1 x 8.0 ml	Streptococcal Extraction Reagent 1
PL.048	1 x 8.0 ml	Streptococcal Extraction Reagent 2
PL.049	1 x 8.0 ml	Streptococcal Extraction Reagent 3

Additional Information


Storage Conditions:	2-8°C
Precautions:	Refer to Instructions for Use

Quality Control Testing

Organism	Reaction*	Result
<i>Streptococcus pyogenes</i> ATCC 19615 (Group A)	Positive	PASS
<i>Streptococcus agalactiae</i> ATCC 12386 (Group B)	Positive	PASS
<i>Streptococcus equisimilis</i> ATCC 12388 (Group C)	Positive	PASS
<i>Enterococcus faecalis</i> ATCC 49477 (Group D)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12392 (Group F)	Positive	PASS
<i>Streptococcus</i> sp. ATCC 12394 (Group G)	Positive	PASS

*When Extraction Reagents 1, 2 and 3 are used in conjunction to extract group specific carbohydrate antigens and tested against homologous latex reagents.

Release

2022 01 27	
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com