



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: ZN CARBOL FUCHSIN	Lot Number: K24968
Product Code: PL.8005	Expiration Date: 2018 09 30

Additional Information

Description: Magenta solution with slight odour.	Concentrate Lot number: 8843 Production date: 2017/10/09	Storage conditions: 15-25°C Precautions: refer to SDS
--	---	--

Quality Control Testing

Challenged with test organisms:
Mycobacterium tuberculosis H37 Rv NCTC 7416
 Streptomyces griseus NCTC 7807
 Cryptosporidia var parvum Wild strain

Test results:
 Conforms to performance specifications when used in standard Ziehl Neelson stain method, giving crisp red staining of Acid/AlcoholFast bacilli with a clean counterstained background.

Comments:
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

2017/10/09	
	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 Texas 78665-1034
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com