



www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name: Grams Staining Kit (Dilute Carbol Fuchsin)	Lot Number: K19219
Product Code: PL.8057/25	Expiration Date: 2015 11

Product Code	Quantity	Product description
PL.7000/25	1 X 250ml	Crystal Violet
PL.7003/25	1 X 250ml	Grams Iodine
PL.7006/25	1 X 250ml	Grams Differentiator
PL.7015/25	1 X 250ml	Dilute Carbol Fuchsin

Additional Information

Description: Violet, brown, clear and magenta solutions.	Concentrate Lot numbers: 6521/6505/6513 Production date: 2014/12/02	Storage conditions: 15-25°C Precautions: refer to MSDS
--	--	---

Quality Control Testing

Challenged with test organisms:
Escherichia coli NCTC 10418
 Oxford *Staphylococcus aureus* NCTC 6571
 Haemolytic *Streptococcus* Group A NCTC 8198

Test results:
 Conforms to performance specifications when used in standard Grams stain method, giving crisp staining of Gram positive and Gram negative bacteria with a clean counterstained background.

Comments:
 We hereby certify that samples of the above Lot have been tested in accordance with current specifications and that the above is a true record of the results obtained.

Release

14/12/02	<i>maun</i>
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 support@pro-lab.com

U.S.A. ☎ 800 522 7740 Fax 800 332 0450
 21 Cypress Blvd., Suite 1070, Round Rock,
 ussupport@pro-lab.com

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 uksupport@pro-lab.com