

QUALITY CONTROL CERTIFICATE

MANNITOL Diagnostic Tablets

REF: 53011
LOT: 2001-1
Date of analysis: 2020-09
Date of expiry: 2024-08
Appearance: Plane tablet, 6 mm in diameter.
Complies.
Average weight: 90 mg.
Complies.

Performance:

Test strains:

E. coli ATCC 25922
Proteus vulgaris ATCC 13315
Staphylococcus aureus ATCC 25923

Product complies with specification.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance