

QUALITY CONTROL CERTIFICATE

HIPPURATE HYDROLYSIS Diagnostic Tablets

REF: 56711
LOT: 1801-1
Date of analysis: 2018-06
Date of expiry: 2022-06
Appearance: Plane tablet, 6 mm in diameter.
Complies.
Average weight: 90 mg.
Complies.

Performance:

Test strains:

Streptococcus agalactiae ATCC 12386
Streptococcus pyogenes ATCC 12344
Staphylococcus aureus 25923
E. coli ATCC 25922

Product complies with specification.

Rosco Diagnostica A/S



Peter Samuelsen
Manager Quality Assurance