

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 927 030-0  
F: +49 30 927 030-30  
info@sifin.de  
www.sifin.de

<b>Product</b>	<b>Anti-Salmonella I (A-E + Vi)</b>
Catalogue no.	TR 1111
Package size	1 ml
<b>Lot no.</b>	<b>118 06 16</b>
<b>Expiry date</b>	<b>2018-06</b>
Intended use	Anti-Salmonella I (A-E) records all <i>Salmonella</i> serovars of the groups A-E and the Vi-antigen in slide agglutination.
Specificity	No agglutination with <i>Salmonella</i> strains of the group F-67 as well as <i>Shigella sonnei</i> and all serotypes of <i>Shigella flexneri</i> .
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 01.08.2016  
place, date

  
signature Head of Quality Control