


Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

Product	Anti-Salmonella I (A-E + Vi)
Catalogue no.	TR1111
Package size	1 ml
Lot no.	6290821
Expiry date	2023-07
Intended use	Anti-Salmonella I (A-E) records all <i>Salmonella</i> serovars of the groups A-E and the Vi-antigen in slide agglutination.
Specificity	No agglutination with <i>Salmonella</i> strains of the group F-67 as well as <i>Shigella sonnei</i> and all serotypes of <i>Shigella flexneri</i> .
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 05.10.21
place, date


signature Head of Quality Control