

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485


sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella OMG O-Group-Pool 60-63, 65-67
Catalogue no.	TR 1174
Package size	1 ml
Lot no.	198 10 14
Expiry date	2016-10
Intended use	Anti-Salmonella OMG records all <i>Salmonella</i> strains belonging to the O-Groups 60-63, 65-67 in slide agglutination.
Specificity	No agglutination with <i>Salmonella</i> strains which are recorded by Anti-Salmonella OMA (O-Group-Pool: A, B, D, E, L), Anti-Salmonella OMB (O-Group-Pool: C, F, G, H), Anti-Salmonella OMC (O-Group-Pool: I, J, K, M, N, O, P), Anti-Salmonella OMD (O-Group-Pool: Q, R, S, T, U, V, W), Anti-Salmonella OME (O-Group-Pool: X, Y, Z, 51, 52, 53) and Anti-Salmonella OMF (O-Group-Pool: 54-59).
Preservatives	Sodium azide 0,9 mg/ml
Storage	2 °C...8 °C

Berlin, 28.11.14
place, date


signature Head of Quality Control