

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

<b>Product</b>	<b>Anti-Salmonella H:a</b>
Catalogue no.	TR 1401
Package size	1 ml
<b>Lot no.</b>	<b>6140719</b>
<b>Expiry date</b>	<b>2022-07</b>
Intended use	Anti-Salmonella H:a records all <i>Salmonella</i> strains, as long as they contain antigen H:a in slide agglutination.
Specificity	No agglutination with heterologous <i>Salmonella</i> H antigens and <i>Salmonella</i> O antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C... 8 °C

Berlin, 09.08.19  
place, date

F. Ludwig  
signature Head of Quality Control