

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella H:z₁₀
Catalogue no.	TR 1427
Package size	1 ml
Lot no.	4030417
Expiry date	2020-04
Intended use	Anti-Salmonella H:z ₁₀ records all <i>Salmonella</i> strains, as long as they contain antigen H:z ₁₀ in slide agglutination.
Specificity	No agglutination with heterologous <i>Salmonella</i> H antigens and <i>Salmonella</i> O antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin,
place, date

26.05.2017

signature Head of Quality Control

