

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella H:z₂₃
Catalogue no.	TR 1440
Package size	1 ml
Lot no.	144 07 14
Expiry date	2018-07
Intended use	Anti-Salmonella H:z ₂₃ records all <i>Salmonella</i> strains, as long as they contain antigen H:z ₂₃ (H:z ₄ ,z ₂₃ ; H:z ₄ ,z ₂₃ ,z ₃₂) in slide agglutination.
Specificity	No agglutination with heterologous <i>Salmonella</i> H antigens and <i>Salmonella</i> O antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 16.09.14
place, date

7. Kuchling
signature Head of Quality Control