

sifin

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Coli O 78:K 80
Catalogue no.	TR 2205
Package size	1 ml
Lot no.	158 08 14
Expiry date	2019-08
Intended use	Anti-Coli O 78:K 80 records all <i>E. coli</i> strains that contain the antigen O 78:K 80 in slide agglutination. In the Widal reaction the test reagent reaches a titre of 1:1600 (tube and MTP) with the homologous <i>E. coli</i> strains; deviations of ± 1 titre step are acceptable.
Specificity	Since unspecific co-agglutinations with other <i>E. coli</i> strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled <i>E. coli</i> isolate in the Widal test.
Preservatives	Sodium azide 0.9 mg/ ml
Storage	2 °C...8 °C

Berlin, 11.11.19
place, date

7. Ludwig
signature Head of Quality Control