

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321
13088 Berlin, Germany

T: +49 30 927 030-0
F: +49 30 927 030-30
info@sifin.de
www.sifin.de

Product	Anti-Salmonella O:22
Catalogue no.	TS 1332
Package size	1 ml
Lot no.	0121119
Expiry date	2022-11
Intended use	Anti-Salmonella O:22 records all Salmonella strains of group G (O:13), as long as they contain antigen O:22.
Specificity	No agglutination with heterologous Salmonella O antigens and Salmonella H antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 09.01.20
place, date

7. Len den
signature Head of Quality Control