

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321  
13088 Berlin, Germany

T: +49 30 700 144-0  
F: +49 30 700 144-30  
info@sifin.de  
www.sifin.de

<b>Product</b>	<b>Anti-Salmonella H:z<sub>4</sub>, z<sub>23</sub></b>
Catalogue no.	TS1425
Package size	1 ml
<b>Lot no.</b>	<b>7871020</b>
<b>Expiry date</b>	<b>2025-09</b>
Intended use	Anti-Salmonella H:z <sub>4</sub> ,z <sub>23</sub> records all <i>Salmonella</i> strains, as long as they contain the antigen combination H:z <sub>4</sub> ,z <sub>23</sub> , H:z <sub>4</sub> ,z <sub>24</sub> , H:z <sub>4</sub> ,z <sub>32</sub> in slide agglutination..
Specificity	No agglutination with heterologous <i>Salmonella</i> H antigens and <i>Salmonella</i> O antigens.
Preservatives	Sodium azide 0.9 mg/ml
Storage	2 °C...8 °C

Berlin, 27. 10. 2020  
\_\_\_\_\_

place, date

*F. Luchj*

\_\_\_\_\_

signature Head of Quality Control